



財團法人全國認證基金會
Taiwan Accreditation Foundation

Certificate of Accreditation

(Certificate No : L4440-260507)

This is to certify that

Super Laboratory Co.,Ltd.
Contract Research Organization

3F. and 4F., No. 24, Wuquan 3rd Rd., Wugu Dist., New Taipei City, Taiwan (R.O.C.)

is accredited in respect of laboratory

Accreditation Criteria : ISO/IEC 17025:2017 ; CNS 17025:2018

Accreditation Number : 4440

Originally Accredited : May 07, 2026

Effective Period : May 07, 2026 to May 06, 2029

Accredited Scope : Testing Field, see described in the Appendix



Scan to verify

Yi-Ling Chen

Yi-Ling Chen
President, Taiwan Accreditation Foundation
May 07, 2026

Accreditation Number : 4440

Laboratory Head : TSAI, Yueh-Ting

▼ 15. 99 Medical Devices

Medical device

B162 Platelet Activation Test

ISO 10993-4 (β-TG Method)

Approval Signatory: CHEN, Yu-Lun

B162 Blood Coagulation Test

ISO 10993-4 (TAT Method)

Approval Signatory: CHEN, Yu-Lun

B300 Acute Systemic Toxicity Test

ISO 10993-11

Approval Signatory: CHEN, Yu-Lun

B300 Subacute Systemic Toxicity Study in Rat

ISO 10993-11

Approval Signatory: CHEN, Yu-Lun

B300 Cytotoxicity Test

ISO 10993-5 (L929, MTT Method)

Approval Signatory: CHEN, Yu-Lun

B310 Rodent Peripheral Blood Micronucleus Test

ISO 10993-3 (ISO/TR 10993-33)

Approval Signatory: CHEN, Yu-Lun

B310 Bacterial Reverse Mutation Test

ISO 10993-3 (ISO/TR 10993-33)

Approval Signatory: CHEN, Yu-Lun

B310 *In vitro* Mammalian Cell Chromosomal Aberration Test

ISO 10993-3 (ISO/TR 10993-33)

Approval Signatory: CHEN, Yu-Lun



B351 Guinea Pig Skin Sensitization Test (Maximization Method)
ISO 10993-10

Approval Signatory: CHEN, Yu-Lun

B352 Rabbit Pyrogen Test
USP-NF <151>

Approval Signatory: CHEN, Yu-Lun

B354 Rabbit Eye Irritation Test
ISO 10993-23

Approval Signatory: CHEN, Yu-Lun

B354 Rabbit Intracutaneous Irritation Test
ISO 10993-23

Approval Signatory: CHEN, Yu-Lun

B354 Rabbit Vaginal Irritation Test
ISO 10993-23

Approval Signatory: CHEN, Yu-Lun

B354 Rabbit Skin Irritation Test
ISO 10993-23

Approval Signatory: CHEN, Yu-Lun

B355 Rabbit Implantation Test (Subcutaneous, Muscal, Bone)
ISO 10993-6

Approval Signatory: CHEN, Yu-Lun

B356 Hemolytic Test
ASTM-F756 (Hemoglobin Lysis)

Approval Signatory: CHEN, Yu-Lun

B357 Endotoxin Test
In-house method: Endotoxin Test - Kinetic-Chromogenic Assay (Document No.: SOPP-325)
Refer to USP-NF <85>, USP-NF <161>, ISO 10993-12
Limit of Detection: 0.005 EU/mL

Approval Signatory: CHIANG, Chien-Yi

(Null below)

P3, total 3 pages

The Appendix forms an integral part of this Certificate, which shall be invalid when use without the Appendix

